

Cyclodienes Plate, Concise ELISA Procedure




BRAXIS

Cyclodienes Plate, Detailed ELISA Procedure




2. Addition of Antibody Solution

Add 100 uL of the antibody solution to the individual wells successively using a multichannel pipette. Cover the wells with parafilm or tape and mix the contents by moving the strip holder in a rapid circular motion on the benchtop. Be careful not to spill contents. Incubate the strips for 60 min at room temperature.



3. Washing of Plates


After incubation, remove the covering and vigorously shake the contents of the wells into a sink. Wash the strips three times with a multi-channel pipette or wash bottle using the diluted 1X washing buffer solution. Please use at least a volume of 250 uL of washing buffer for each well and each washing step. Remaining buffer in the wells should be removed by patting the plate dry on a stack of paper towels.



OR




Add 100 uL of the enzyme conjugate to the individual wells successively using a multi- channel pipette or a stepping pipette. Incubate the strips for 30 minutes at room temperature.




5. Washing of Plates

After incubation, remove the covering and vigorously shake the contents of the wells into a sink. Wash the strips four times with a multi-channel pipette or wash bottle using the diluted 1X washing buffer solution. Please use at least a volume of 250 uL of washing buffer for each well and each washing step. Remaining buffer in the wells should be removed by patting the plate dry on a stack of paper towels.




6. Addition of Substrate/Color Solution

Add 100 uL of substrate/color solution to the individual wells successively using a multichannel pipette or a stepping pipette. Cover the wells with parafilm or tape and mix the contents by moving the strip holder in a rapid circular motion on the benchtop. Be careful not to spill contents. Incubate the strips for 20 min at room temperature.




I. Addition of Stopping Solution

Add 50 uL of stop solution to the wells in the same sequence as for the substrate solution using a multi- channel pipette or a stepping pipette.



8. Measurement of Color

Read the absorbance at 450 nm using a microplate ELISA reader. Calculate results.



For Ordering or Technical Assistance Contact: ABRAXIS, LLC 54 Steamwhistle Drive, Warminster, PA 18974 Phone: 215-357-3911 Fax: 215-357-5232 www.abraxiskits.com

